

Evaluation of Dentist's Satisfaction of Marginal Fit in Fixed Partial Dentures Fabricated Using Cad/Cam Systems - An Institution Based Retrospective Study

Type of Manuscript : Retrospective Study

Running Title : Dentist's satisfaction of marginal fit in FPD using CAD/CAM systems

Dr. Harini Sri

Saveetha Dental College
Saveetha Institute of Medical and Technical Sciences (SIMATS)
Saveetha University
Chennai
Email -151909005.sdc@saveetha.com

Dr. Nabeel Ahmed

Senior lecturer
Department of Prosthodontics
Saveetha Dental College
Saveetha Institute of Medical and Technical Sciences (SIMATS)
Saveetha University
Chennai
Email -nabeeln.sdc@saveetha.com

Dr. Keerthi Shashanka

Senior lecturer
Department of Prosthodontics
Saveetha Dental College
Saveetha Institute of Medical and Technical Sciences (SIMATS)
Saveetha University
Chennai
Email - keerthis.sdc@saveetha.com

Corresponding Author:

Dr. Harini Sri
Saveetha Dental College
Saveetha Institute of Medical and Technical Sciences (SIMATS)
Saveetha University
Chennai
Email - 151909005.sdc@saveetha.com

ABSTRACT:

Accuracy and quality of CAD CAM in producing fixed partial dentures, crowns and bridges is acceptable and improving every day. Marginal adaptation is one of the most important factors for success and clinical longevity of the treatment. The aim of this study was to retrospectively evaluate the dentist's satisfaction of marginal fit in fixed partial dentures fabricated using

CAD/CAM systems. The data for this study was collected from June 2019 to March 2020 at a private dental institution in Chennai. A total of 507 patients who received fixed partial dentures were included in the study. Data regarding the marginal fit in relation to each sextant was collected from the case record. Chi-square test was applied to see associations of marginal fit satisfaction in each sextant. The results showed that 94.1% (n=477) of the fixed partial dentures a satisfactory marginal fit, while the marginal fit in 5.9% (n=30) of the FPDs were non satisfactory. Fixed partial dentures in sextant 2 showed 30% (n=9) non satisfaction followed by sextant 1 with 20% (n=6). The findings of this study shows that the FPDs fabricated using CAD/CAM systems had a satisfactory marginal fit and comparatively sextant 2 presented with high marginal misfit than the other sextants.

Key words: CAD/CAM system; Dentist's satisfaction, Fixed partial denture; Marginal fit.

INTRODUCTION:

Nowadays, patients are well aware of his or her treatment options, and CAD/CAM technology may help the dentist meet such a patient's needs. With access to well designed software, attention to user-friendliness and positive clinical performance history, CAD/CAM systems are a significant stride forward into advanced dental treatment (Trost, Stines and Burt, 2006). Accuracy and quality of CAD CAM in producing fixed partial dentures, crowns and bridges is acceptable and improving day by day (Jacob *et al.*, 2016). This helps the clinicians to maintain total product and artistic control of the restorations to be fabricated and seated. The designing softwares deliver a product that may need only endpoint characterization, staining or glazing. The computer and milling processes diminish the potential inaccuracies that are usually seen in lost wax technique and are able to provide a restoration that fits within the 50-micrometer range established by the American Dental Association (Estafan *et al.*, 2003; Abdullah, Tsitrou and Pollington, 2016).

Fixed partial denture is a common practise over ages for replacing missing teeth. Marginal adaptation is one of the most important factors for the success and clinical longevity of Fixed partial dentures. Inadequate fit leads to accumulation of plaque, which increases the risk of periodontal diseases (Björn, Björn and Grkovic, 1970). Misfit could contribute to cement dissolution resulting in caries and endodontic inflammation (Cooper *et al.*, 1971). Some clinicians remain reluctant to incorporate CAD/CAM technology into their practices despite its increasing popularity. One of the main concerns relates to the size of the marginal gap that occurred in many of the restorations produced by the early CEREC systems. The fit of earlier versions of CAD/CAM restorations has been reported by some authors and clinicians (Hickel *et al.*, 1998). Some clinicians have even associated the poor fit of older systems with recently introduced products when, in fact, marginal adaptation has been vastly improved and can exceed that of conventional laboratory methods (Castillo-de-Oyague *et al.*, 2012). Modern CAD/ CAM technology has been shown to allow the fabrication of better fitting partial fixed dental prostheses and implant frameworks than those made with the traditional techniques (Lopez-Suarez *et al.*, 2016). Studies agree that the fit of the CAD/CAM restoration rivals that of conventionally fabricated restorations.

Different methods have been used to assess marginal and internal adaptation of indirect restorations. Clinically, marginal fit can be estimated directly using a probe with known dimensions, a dental mirror, and adequate illumination and magnification. Marginal fit can also

be indirectly evaluated radiographically, and through epoxy resin replicas by light and scanning electron microscopy (An *et al.*, 2014). There is no consensus on which is the best method to evaluate marginal and internal adaptation of indirect restorations. There are various studies comparing marginal fit of FPDs fabricated using different techniques and materials. But there are very few studies that analyse dentist's satisfaction of marginal fit in FPDs in different sextant. Our team has rich experience in research and we have collaborated with numerous authors over various topics in the past decade (Subramanyam *et al.*, 2018)('Fluoride, fluoridated toothpaste efficacy and its safety in children - review', 2018; Ezhilarasan, 2018; Felicita, 2018; Kavarthapu and Thamaraiselvan, 2018; Krishnan *et al.*, 2018; Marimuthu *et al.*, 2018; Nair *et al.*, 2018; Padavala and Sukumaran, 2018; Pandian, Krishnan and Kumar, 2018; Rajeshkumar *et al.*, 2018; Rao and Kumar, 2018; Vijayashree Priyadharsini, Smiline Girija and Paramasivam, 2018; Abhinav *et al.*, 2019; Ke *et al.*, 2019; Mehta *et al.*, 2019; Panchal, Jeevanandan and Subramanian, 2019; Ponnulakshmi *et al.*, 2019; Ramesh *et al.*, 2019; Sridharan *et al.*, 2019; Sweta, Abhinav and Ramesh, 2019; Wu *et al.*, 2019; Palati *et al.*, 2020; Paramasivam, Vijayashree Priyadharsini and Raghunandhakumar, 2020).

The objective of this study was to retrospectively evaluate the dentist's satisfaction of marginal fit in fixed partial dentures fabricated using CAD/CAM systems .

MATERIALS AND METHODS:

This retrospective study was conducted from June 2019 to March 2020, in the Department of Prosthodontics, in a private dental institution in Chennai, after taking ethical clearance from the Institutional Review Committee. Case records of patients who underwent Fixed partial denture treatment were collected and the missing entries were omitted (Purposive sampling). So a total of 507 case records were obtained. Written informed consent was taken from each patient participating in the study. Case sheet of each patient included age, chief complaint, history of presenting illness, relevant medical and dental history, intraoral and extraoral examination, treatment plan, preoperative and post-operative radiographs and photographs and an elaborate summary of the treatment procedure done on the patient. It had a separate column for the dentist's opinion on the treatment. So the data on the marginal fit of FPDs were collected. Photographs and radiographs were used for cross verification of data to reduce errors. To minimise the sampling bias, the study was double blinded by an analyser and a reviewer.

After well refined tooth preparation, gingival retraction was done using Ultrapak™ knitted cord. Double cords were used during cord packing. Double stage putty wash impressions were taken using Virtual®Putty and Ivoclar Vivadent Virtual Refill Light Body Regular Set VPS Impression Material. The master cast of all the patients were scanned using the Medit T series extraoral scanner, and the framework was designed in 3 shape software. The STL file was then imported to I MES I CORE milling machine and sintered.

The collected data was entered in Microsoft Excel 2010 and a specific statistical software, IBM SPSS 19.0 version was used to analyse the data. Descriptive statistics was performed for each group and Chi-square test was performed to see the association between satisfaction among different sextants.

RESULTS:

A total of 507 case reports were examined among whom 298 (58.8%) were male and 209

(41.2%) were female aged between 20-60 years (figure 1). The results showed that 93.88% (n=476) of the fixed partial dentures fabricated using CAD/CAM system had a satisfactory marginal fit, while the marginal fit in 6.11% (n=31) of the FPDs were non satisfactory (figure 2). Chi square test was used to determine the association of marginal fit satisfaction and sextant and was found to be statistically significant (p Value = 0.01) was obtained. The study reported that the marginal fit was highly satisfactory in sextant 6 and sextant 5 with 25.0% and 24.79% respectively, followed by sextant 4 with 19.33%, sextant 3 with 15.76%, sextant 1 with 11.76% and 3.36% in sextant 2. It was observed that sextant 2 had a non-satisfactory fit of 32.26% followed by sextant 1 with 19.35%, sextant 5 with 16.13%, sextant 3 with 12.9%, sextant 6 and 4 with 9.77% (figure 3, table 1).

DISCUSSION:

Previously our team has conducted numerous original studies over the past 5 years (Ashok *et al.*, 2014; Venugopalan *et al.*, 2014; Ashok and Suvitha, 2016; Ganapathy *et al.*, 2016; Selvan and Ganapathy, 2016; Subasree, Murthykumar and Dhanraj, 2016; Vijayalakshmi and Ganapathy, 2016; Ajay *et al.*, 2017; Ganapathy, Kannan and Venugopalan, 2017; Jyothi *et al.*, 2017; Ranganathan, Ganapathy and Jain, 2017; Ariga *et al.*, 2018; Basha, Ganapathy and Venugopalan, 2018; Kannan and Venugopalan, 2018; Duraisamy *et al.*, 2019). The idea for this study stemmed from the current interest in our community. Marginal fit is an essential aspect that determines the success of a fixed partial prosthesis. Failure of FPD bridges usually occur due to improper case selection, faulty diagnosis and treatment plan, inaccurate clinical and laboratory procedures and poor patient care and maintenance following insertion (Nallaswamy, 2017). This study retrospectively evaluated the dentist's satisfaction of marginal fit in fixed partial dentures fabricated using CAD/CAM systems. The results showed a statistically significant correlation between the marginal fit of FPDs and dentist's satisfaction. According to the results, about 93.8% of the bridges showed satisfactory marginal fit and very meager numbers showed non satisfactory marginal fit. This is in accordance to the results found by various studies that reported that the marginal fit of CAD/CAM restoration is within the limits considered clinically acceptable by ADA specification #8, therefore milled CAD/CAM crowns can be considered a good alternative to more traditional waxing-investing-casting technique (Witkowski, Komine and Gerds, 2006; Miyazaki and Hotta, 2011; Ghoul *et al.*, 2020). The potential reason for this non satisfactory fit has to do with the diamond rotary cutting instruments that the current CAD/CAM mills use and their inability to replicate the features caused by these preparation errors. Furthermore, Renne *et al.* emphasised that, when milling the restorations, variables within the mill itself, such as diamond rotary cutting instrument wear and water quality, may affect the quality of the restoration (Renne *et al.*, 2012). It should be noted that minor errors in designing softwares and scanners can also add on to the marginal misfit.

The results of the present study reported that dentists found satisfactory marginal fit of the FPDs in sextant 5 and sextant 6. Sadatullah *et al.* reported that most of the dentists considered maxillary teeth preparation to be more difficult than mandibular teeth reduction (Syed, Al-Moaleem and Shariff, 2016). Maximum misfit was observed in sextant 2 followed by sextant 1. This could probably be because of the curvature of the arch in the upper anterior region. But Pound *et al.*, reported that achieving success in the lower anterior bridge was more difficult (Pound, 1955). Nowadays there is an increase in the number of laboratories using CAD/CAM systems to fabricate fixed partial dentures and may also have similar issues with poor

preparations. Errors in preparation design, particularly involving the margin, are easier to cope with when a laboratory uses the lost wax technique to fabricate alloy restorations (Renne *et al.*, 2012). However it is worth noting that all preparations were thoroughly examined by experienced clinicians. However, it is the hope of the authors that clinicians will be able to recognize and correct these common preparation errors in order to improve the overall fit of their restorations.

Since this study was conducted in an institution-based set-up and the sample size is small, this might bias the study results, as the selected sample is not representative. The inclusion and exclusion criteria were not well specified and the abutment conditions were not evaluated. Hence, further studies are required to explore about the marginal fit of CAD/CAM milled restorations. Our institution is passionate about high quality evidence based research and has excelled in various fields ((Pc, Marimuthu and Devadoss, 2018; Ramesh *et al.*, 2018; Vijayashree Priyadharsini, Smiline Girija and Paramasivam, 2018; Ezhilarasan, Apoorva and Ashok Vardhan, 2019; Ramadurai *et al.*, 2019; Sridharan *et al.*, 2019; Vijayashree Priyadharsini, 2019; Chandrasekar *et al.*, 2020; Mathew *et al.*, 2020; R *et al.*, 2020; Samuel, 2021)

"The ideal CAD / CAM system" for many years is a dream of many researchers. The data observed from this retrospective study provides the clinicians with the information regarding the performance of CAD/CAM systems in marginal adaptation.

CONCLUSION:

Within the limitations of this study it was observed that the marginal fit of CAD/CAM milled fixed partial dentures were satisfactory. Among the sextants, satisfactory marginal fit was reported in sextant 5 and 6, maximum marginal discrepancy was seen in sextant 2 followed by sextant 1.

AUTHOR CONTRIBUTIONS:

Dr. Harini Sri

Author contributed in the conception, design, acquisition of data, analysis, drafting the article and interpretation of data.

ACKNOWLEDGEMENT

We would like to acknowledge Saveetha Dental College and Hospital for providing complete patient details required for the study purpose and their constant help and support for this research.

CONFLICT OF INTEREST

The authors declare no conflict of interest.

REFERENCE:

[1]. Abdullah, A. O., Tsitrou, E. A. and Pollington, S. (2016) 'Comparative in vitro evaluation of CAD/CAM vs conventional provisional crowns', *Journal of applied oral science: revista FOB*, 24(3), pp. 258–263.

- [2]. Abhinav, R. P. *et al.* (2019) 'The Patterns and Etiology of Maxillofacial Trauma in South India', *Annals of maxillofacial surgery*, 9(1), pp. 114–117.
- [3]. Ajay, R. *et al.* (2017) 'Effect of surface modifications on the retention of cement-retained implant crowns under fatigue loads: An In vitro study', *Journal of Pharmacy And Bioallied Sciences*, p. 154. doi: 10.4103/jpbs.jpbs_146_17.
- [4]. An, S. *et al.* (2014) 'Evaluating the marginal fit of zirconia copings with digital impressions with an intraoral digital scanner', *The Journal of Prosthetic Dentistry*, pp. 1171–1175. doi: 10.1016/j.prosdent.2013.12.024.
- [5]. Ariga, P. *et al.* (2018) 'Determination of Correlation of Width of Maxillary Anterior Teeth using Extraoral and Intraoral Factors in Indian Population: A Systematic Review', *World Journal of Dentistry*, pp. 68–75. doi: 10.5005/jp-journals-10015-1509.
- [6]. Ashok, V. *et al.* (2014) 'Lip Bumper Prosthesis for an Acromegaly Patient: A Clinical Report', *Journal of Indian Prosthodontic Society*, 14(Suppl 1), pp. 279–282.
- [7]. Ashok, V. and Suvitha, S. (2016) 'Awareness of all ceramic restoration in rural population', *Research Journal of Pharmacy and Technology*, p. 1691. doi: 10.5958/0974-360x.2016.00340.1.
- [8]. Basha, F. Y. S., Ganapathy, D. and Venugopalan, S. (2018) 'Oral Hygiene Status among Pregnant Women', *Research Journal of Pharmacy and Technology*, p. 3099. doi: 10.5958/0974-360x.2018.00569.3.
- [9]. Björn, A. L., Björn, H. and Grkovic, B. (1970) 'Marginal fit of restorations and its relation to periodontal bone level. II. Crowns', *Odontologisk revy*, 21(3), pp. 337–346.
- [10]. Castillo-de-Oyague, R. *et al.* (2012) 'Vertical misfit of laser-sintered and vacuum-cast implant-supported crown copings luted with definitive and temporary luting agents', *Medicina Oral Patología Oral y Cirugía Bucal*, pp. e610–e617. doi: 10.4317/medoral.17997.
- [11]. Chandrasekar, R. *et al.* (2020) 'Development and validation of a formula for objective assessment of cervical vertebral bone age', *Progress in orthodontics*, 21(1), p. 38.
- [12]. Cooper, T. M. *et al.* (1971) 'Effect of venting on cast gold full crowns', *The Journal of prosthetic dentistry*, 26(6), pp. 621–626.
- [13]. Duraisamy, R. *et al.* (2019) 'Compatibility of Nonoriginal Abutments With Implants: Evaluation of Microgap at the Implant-Abutment Interface, With Original and Nonoriginal Abutments', *Implant dentistry*, 28(3), pp. 289–295.
- [14]. Estafan, D. *et al.* (2003) 'Scanning electron microscope evaluation of CEREC II and CEREC III inlays', *General dentistry*, 51(5), pp. 450–454.
- [15]. Ezhilarasan, D. (2018) 'Oxidative stress is bane in chronic liver diseases: Clinical and experimental perspective', *Arab journal of gastroenterology: the official publication of the Pan-Arab Association of Gastroenterology*, 19(2), pp. 56–64.
- [16]. Ezhilarasan, D., Apoorva, V. S. and Ashok Vardhan, N. (2019) 'Syzygium cumini extract induced reactive oxygen species-mediated apoptosis in human oral squamous carcinoma cells', *Journal of oral pathology & medicine: official publication of the International Association of Oral Pathologists and the American Academy of Oral Pathology*, 48(2), pp. 115–121.
- [17]. Felicita, A. S. (2018) 'Orthodontic extrusion of Ellis Class VIII fracture of maxillary lateral incisor - The sling shot method', *The Saudi dental journal*, 30(3), pp. 265–269.
- [18]. 'Fluoride, fluoridated toothpaste efficacy and its safety in children - review' (2018) *International journal of pharmaceutical research*, 10(04). doi: 10.31838/ijpr/2018.10.04.017.
- [19]. Ganapathy, D. *et al.* (2016) 'Effect of Resin Bonded Luting Agents Influencing

Marginal Discrepancy in All Ceramic Complete Veneer Crowns', *Journal of clinical and diagnostic research: JCDR*, 10(12), pp. ZC67–ZC70.

[20]. Ganapathy, D. M., Kannan, A. and Venugopalan, S. (2017) 'Effect of Coated Surfaces influencing Screw Loosening in Implants: A Systematic Review and Meta-analysis', *World Journal of Dentistry*, pp. 496–502. doi: 10.5005/jp-journals-10015-1493.

[21]. Ghoul, W. A. E. et al. (2020) 'Effect of different CAD-CAM materials on the marginal and internal adaptation of endocrown restorations: An in vitro study', *The Journal of Prosthetic Dentistry*, pp. 128–134. doi: 10.1016/j.prosdent.2018.10.024.

[22]. Hickel, R. et al. (1998) 'New direct restorative materials. FDI Commission Project', *International dental journal*, 48(1), pp. 3–16.

[23]. Jacob, S. A. et al. (2016) 'CAD CAM - Understanding the Basics: A Review', *Journal of Dentistry and Oral Care*, pp. 1–6. doi: 10.15436/2379-1705.16.1046.

[24]. Jyothi, S. et al. (2017) 'Periodontal Health Status of Three Different Groups Wearing Temporary Partial Denture', *Research Journal of Pharmacy and Technology*, p. 4339. doi: 10.5958/0974-360x.2017.00795.8.

[25]. Kannan, A. and Venugopalan, S. (2018) 'A systematic review on the effect of use of impregnated retraction cords on gingiva', *Research Journal of Pharmacy and Technology*, p. 2121. doi: 10.5958/0974-360x.2018.00393.1.

[26]. Kavarthapu, A. and Thamaraiselvan, M. (2018) 'Assessing the variation in course and position of inferior alveolar nerve among south Indian population: A cone beam computed tomographic study', *Indian journal of dental research: official publication of Indian Society for Dental Research*, 29(4), pp. 405–409.

[27]. Ke, Y. et al. (2019) 'Photosynthesized gold nanoparticles from *Catharanthus roseus* induces caspase-mediated apoptosis in cervical cancer cells (HeLa)', *Artificial cells, nanomedicine, and biotechnology*, 47(1), pp. 1938–1946.

[28]. Krishnan, R. P. et al. (2018) 'Surgical Specimen Handover from Operation Theater to Laboratory: A Survey', *Annals of maxillofacial surgery*, 8(2), pp. 234–238.

[29]. Lopez-Suarez, C. et al. (2016) 'Marginal Vertical Discrepancies of Monolithic and Veneered Zirconia and Metal-Ceramic Three-Unit Posterior Fixed Dental Prostheses', *The International Journal of Prosthodontics*, pp. 256–258. doi: 10.11607/ijp.4541.

[30]. Marimuthu, M. et al. (2018) 'Canonical Wnt pathway gene expression and their clinical correlation in oral squamous cell carcinoma', *Indian journal of dental research: official publication of Indian Society for Dental Research*, 29(3), pp. 291–297.

[31]. Mathew, M. G. et al. (2020) 'Evaluation of adhesion of *Streptococcus mutans*, plaque accumulation on zirconia and stainless steel crowns, and surrounding gingival inflammation in primary molars: Randomized controlled trial', *Clinical oral investigations*, pp. 1–6.

[32]. Mehta, M. et al. (2019) 'Oligonucleotide therapy: An emerging focus area for drug delivery in chronic inflammatory respiratory diseases', *Chemico-biological interactions*, 308, pp. 206–215.

[33]. Miyazaki, T. and Hotta, Y. (2011) 'CAD/CAM systems available for the fabrication of crown and bridge restorations', *Australian dental journal*, 56 Suppl 1, pp. 97–106.

[34]. Nair, M. et al. (2018) 'Comparative evaluation of post-operative pain after pulpectomy with k-files, kedo-s files and mtwo files in deciduous molars -a randomized clinical trial', *Brazilian dental science*, 21(4), p. 411.

[35]. Nallaswamy, D. (2017) *Textbook of Prosthodontics*. JP Medical Ltd.

[36]. Padavala, S. and Sukumaran, G. (2018) 'Molar Incisor Hypomineralization and Its

Prevalence', *Contemporary clinical dentistry*, 9(Suppl 2), pp. S246–S250.

- [37]. Palati, S. *et al.* (2020) 'Knowledge, Attitude and practice survey on the perspective of oral lesions and dental health in geriatric patients residing in old age homes', *Indian journal of dental research: official publication of Indian Society for Dental Research*, 31(1), pp. 22–25.
- [38]. Panchal, V., Jeevanandan, G. and Subramanian, E. (2019) 'Comparison of instrumentation time and obturation quality between hand K-file, H-files, and rotary Kedo-S in root canal treatment of primary teeth: A randomized controlled trial', *Journal of the Indian Society of Pedodontics and Preventive Dentistry*, 37(1), pp. 75–79.
- [39]. Pandian, K. S., Krishnan, S. and Kumar, S. A. (2018) 'Angular photogrammetric analysis of the soft-tissue facial profile of Indian adults', *Indian journal of dental research: official publication of Indian Society for Dental Research*, 29(2), pp. 137–143.
- [40]. Paramasivam, A., Vijayashree Priyadharsini, J. and Raghunandhakumar, S. (2020) 'N6-adenosine methylation (m6A): a promising new molecular target in hypertension and cardiovascular diseases', *Hypertension research: official journal of the Japanese Society of Hypertension*, 43(2), pp. 153–154.
- [41]. Pc, J., Marimuthu, T. and Devadoss, P. (2018) 'Prevalence and measurement of anterior loop of the mandibular canal using CBCT: A cross sectional study', *Clinical implant dentistry and related research*. Available at: <https://europepmc.org/article/med/29624863>.
- [42]. Ponnulakshmi, R. *et al.* (2019) 'In silico and in vivo analysis to identify the antidiabetic activity of beta sitosterol in adipose tissue of high fat diet and sucrose induced type-2 diabetic experimental rats', *Toxicology mechanisms and methods*, 29(4), pp. 276–290.
- [43]. Pound, E. (1955) 'The problem of the lower anterior bridge', *The Journal of Prosthetic Dentistry*, pp. 543–545. doi: 10.1016/0022-3913(55)90183-9.
- [44]. Rajeshkumar, S. *et al.* (2018) 'Biosynthesis of zinc oxide nanoparticles using *Mangifera indica* leaves and evaluation of their antioxidant and cytotoxic properties in lung cancer (A549) cells', *Enzyme and microbial technology*, 117, pp. 91–95.
- [45]. Ramadurai, N. *et al.* (2019) 'Effectiveness of 2% Articaine as an anesthetic agent in children: randomized controlled trial', *Clinical oral investigations*, 23(9), pp. 3543–3550.
- [46]. Ramesh, A. *et al.* (2018) 'Comparative estimation of sulfiredoxin levels between chronic periodontitis and healthy patients - A case-control study', *Journal of periodontology*, 89(10), pp. 1241–1248.
- [47]. Ramesh, A. *et al.* (2019) 'Esthetic lip repositioning: A cosmetic approach for correction of gummy smile - A case series', *Journal of Indian Society of Periodontology*, 23(3), pp. 290–294.
- [48]. Ranganathan, H., Ganapathy, D. M. and Jain, A. R. (2017) 'Cervical and Incisal Marginal Discrepancy in Ceramic Laminate Veneering Materials: A SEM Analysis', *Contemporary clinical dentistry*, 8(2), pp. 272–278.
- [49]. Rao, T. D. and Kumar, M. P. S. (2018) 'Analgesic efficacy of paracetamol vs ketorolac after dental extractions', *Journal of advanced pharmaceutical technology & research*, 11(8), p. 3375.
- [50]. Renne, W. *et al.* (2012) 'Predicting marginal fit of CAD/CAM crowns based on the presence or absence of common preparation errors', *The Journal of prosthetic dentistry*, 108(5), pp. 310–315.
- [51]. R, H. *et al.* (2020) 'CYP2 C9 polymorphism among patients with oral squamous cell carcinoma and its role in altering the metabolism of benzo[a]pyrene', *Oral Surgery, Oral Medicine, Oral Pathology and Oral Radiology*, pp. 306–312. doi: 10.1016/j.oooo.2020.06.021.

- [52]. Samuel, S. R. (2021) 'Can 5-year-olds sensibly self-report the impact of developmental enamel defects on their quality of life?', *International journal of paediatric dentistry / the British Paedodontic Society [and] the International Association of Dentistry for Children*, 31(2), pp. 285–286.
- [53]. Selvan, S. R. and Ganapathy, D. (2016) 'Efficacy of fifth generation cephalosporins against methicillin-resistant Staphylococcus aureus-A review', *Research Journal of Pharmacy and Technology*, p. 1815. doi: 10.5958/0974-360x.2016.00369.3.
- [54]. Sridharan, G. et al. (2019) 'Evaluation of salivary metabolomics in oral leukoplakia and oral squamous cell carcinoma', *Journal of oral pathology & medicine: official publication of the International Association of Oral Pathologists and the American Academy of Oral Pathology*, 48(4), pp. 299–306.
- [55]. Subasree, S., Murthykumar, K. and Dhanraj (2016) 'Effect of Aloe Vera in Oral Health-A Review', *Research Journal of Pharmacy and Technology*, p. 609. doi: 10.5958/0974-360x.2016.00116.5.
- [56]. Subramanyam, D. et al. (2018) 'Comparative evaluation of salivary malondialdehyde levels as a marker of lipid peroxidation in early childhood caries', *European journal of dentistry*, 12(1), pp. 67–70.
- [57]. Sweta, V. R., Abhinav, R. P. and Ramesh, A. (2019) 'Role of virtual reality in pain perception of patients following the administration of local anesthesia', *Annals of maxillofacial surgery*, 9(1), pp. 110–113.
- [58]. Syed, S., Al-Moaleem, M. M. and Shariff, M. (2016) 'The quality assessment of teeth prepared by fresh graduates for ceramo-metal full coverage crowns', *The Saudi Journal for Dental Research*, pp. 24–28. doi: 10.1016/j.sjdr.2015.01.003.
- [59]. Trost, L., Stines, S. and Burt, L. (2006) 'Making informed decisions about incorporating a CAD/CAM system into dental practice', *The Journal of the American Dental Association*, p. 32S–36S. doi: 10.14219/jada.archive.2006.0399.
- [60]. Venugopalan, S. et al. (2014) 'Magnetically retained silicone facial prosthesis', *Nigerian journal of clinical practice*, 17(2), pp. 260–264.
- [61]. Vijayalakshmi, B. and Ganapathy, D. (2016) 'Medical management of cellulitis', *Research Journal of Pharmacy and Technology*, p. 2067. doi: 10.5958/0974-360x.2016.00422.4.
- [62]. Vijayashree Priyadharsini, J. (2019) 'In silico validation of the non-antibiotic drugs acetaminophen and ibuprofen as antibacterial agents against red complex pathogens', *Journal of periodontology*, 90(12), pp. 1441–1448.
- [63]. Vijayashree Priyadharsini, J., Smiline Girija, A. S. and Paramasivam, A. (2018) 'In silico analysis of virulence genes in an emerging dental pathogen A. baumannii and related species', *Archives of oral biology*, 94, pp. 93–98.
- [64]. Witkowski, S., Komine, F. and Gerds, T. (2006) 'Marginal accuracy of titanium copings fabricated by casting and CAD/CAM techniques', *The Journal of prosthetic dentistry*, 96(1), pp. 47–52.
- [65]. Wu, F. et al. (2019) 'Biologically synthesized green gold nanoparticles from induce growth-inhibitory effect on melanoma cells (B16)', *Artificial cells, nanomedicine, and biotechnology*, 47(1), pp. 3297–3305.

GRAPHS:




Figure 1: Gender distribution in the study in percentage. Brown colour represents male and purple represents female participation in the study in percentage. 41.2% of female (n=209) and 53.78% (n=298) of male participated in the study.




Figure 2: Frequency of satisfaction of marginal fit. X - axis depicts satisfaction of marginal fit among dentists and Y - axis shows the percentage of Fixed Partial Denture. It was observed that 93.88% (n=476) of the fixed partial dentures fabricated using CAD/CAM system had a satisfactory marginal fit (blue), while the marginal fit in 6.11% (n=31) of the FPDs were non satisfactory (green).




Figure 3 : Satisfaction of marginal fit in each sextant. X - axis depicts the sextant and Y axis represents the percentage of satisfaction of marginal fit. It is observed that the marginal fit was highly satisfactory (blue) in sextant 5 and sextant 6 with 24.79% and 25.0%. Sextant 2 had a non-satisfactory fit (green) of 32.26% followed by sextant 1 with 19.35%. Chi square test was used to determine the association of marginal fit satisfaction and sextant and was found to be statistically significant (p value = 0.01) was obtained.

TABLES:

SEXTANT	MARGINAL FIT				TOTAL	
	SATISFACTORY		NON SATISFACTORY			
	COUNT	%	COUNT	%	COUNT	%
SEXTANT 1	56	11.8%	6	19.4%	62	12.2%
SEXTANT 2	16	3.4%	10	32.3%	26	5.1%
SEXTANT 3	75	15.8%	4	12.9%	79	15.6%
SEXTANT 4	92	19.3%	3	9.7%	95	18.7%
SEXTANT 5	118	24.8%	5	16.1%	123	24.3%
SEXTANT 6	119	25%	3	9.7%	122	24.1%

Table 1: Satisfaction levels in each sextant. Marginal fit was satisfactory in sextant 5 and sextant 6 with 24.79% and 25.0%. Sextant 2 had a non-satisfactory fit of 32.26% followed by sextant 1 with 19.35%. Chi square test was used to determine the association of marginal fit satisfaction and sextant and was found to be statistically significant (p value = 0.01) was obtained.